

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	31803-5K	2018-08-07
Low	31803-51	2018-08-09
Normal	31803-52	2018-08-07
High	31803-53	2018-08-07

MEDONIC

CA620/530

		CONTROL	L
		LOT	31803-51
		Mean ²	± Range ³
RBC	10 ¹² /L	2,17	0,12
MCV	fL	65,9	7,0
RDW	%	13,2	4,5
Hct	%	14,3	3,0
Plt	10 ⁹ /L	95	15
MPV	fL	9,6	1,5
WBC	10 ⁹ /L	3,4	0,3
Hgb	g/dL	5,5	0,3
MCH	pg	25,3	2,0
MCHC	g/dL	38,5	4,0
Lymph#	10 ⁹ /L	1,6	0,4
Mid#	10 ⁹ /L	0,2	0,3
Gran#	10 ⁹ /L	1,6	0,8
Lymph	%	46,7	7,0
Mid	%	5,7	5,0
Gran	%	47,6	7,0

		CONTROL	N
		LOT	31803-52
		Mean ²	± Range ³
		4,03	0,18
		79,8	7,0
		12,4	4,5
		32,2	3,5
		274	30
		9,8	1,5
		9,3	0,6
		11,6	0,4
		28,8	2,0
		36,0	4,0
		4,7	0,6
		0,6	0,5
		4,0	1,0
		50,8	5,0
		6,5	5,0
		42,7	5,0

		CONTROL	H
		LOT	31803-53
		Mean ²	± Range ³
		4,76	0,22
		87,2	7,0
		12,4	4,5
		41,5	4,0
		585	60
		10,2	1,5
		22,2	1,8
		14,9	0,5
		31,3	2,0
		35,9	4,0
		14,3	1,5
		1,3	1,0
		6,6	2,5
		64,4	5,0
		6,0	5,0
		29,6	5,0

		Medonic M-series/M32	Mean ²	± Range ³
		RBC	10 ¹² /L	2,16
MCV	fL	67,4	7,0	
RDW	%	13,2	4,5	
Hct	%	14,6	3,0	
Plt	10 ⁹ /L	85	15	
MPV	fL	9,8	1,5	
WBC	10 ⁹ /L	3,3	0,3	
Hgb	g/dL	5,4	0,3	
MCH	pg	25,0	2,0	
MCHC	g/dL	37,0	4,0	
Lymph#	10 ⁹ /L	1,5	0,4	
Mid#	10 ⁹ /L	0,2	0,3	
Gran#	10 ⁹ /L	1,6	0,8	
Lymph	%	45,4	7,0	
Mid	%	5,4	5,0	
Gran	%	49,2	7,0	

		Mean ²	± Range ³
		82,1	7,0
		12,3	4,5
		33,2	3,5
		250	30
		9,8	1,5
		8,9	0,6
		11,7	0,4
		29,0	2,0
		35,2	4,0
		4,4	0,6
		0,5	0,5
		4,0	1,0
		49,1	5,0
		6,1	5,0
		44,8	5,0

		Mean ²	± Range ³
		90,6	7,0
		12,1	4,5
		43,7	4,0
		536	60
		10,0	1,5
		21,5	1,8
		15,2	0,5
		31,5	2,0
		34,8	4,0
		13,1	1,5
		1,5	1,0
		6,9	2,5
		61,0	5,0
		7,1	5,0
		31,9	5,0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dana (hr)
 - cs Stabilitnosta otevrené ampulky 14 dní (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jour (fr)
 - de Stabilit t ge ffneter Flaschen 14 tage (de)
 - el Διατ ρηση δειγματος μετ  την αποσφραγίση 14 ημερ  (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no  pen r r/glass stabilitet 14 dager (no)
 - pl Trwa o c po otworzeniu fiołki 14 Liczba dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekcnej liekovky 14 dni (sk)
 - sl Stabilita otvorene injekcne liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 d as (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Sredno
 - Sredstvo
 - Pr m r
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - M so
 - M dia
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Srednia
 - M dia
 - M die
 - Srednee
 - Srednja vrednost
 - Priemer
 - Povpre no
 - M dia
 - Medelv rde
 - Ortalama

- 3
- Ob vat
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο ε ρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Arah 



<http://www.medonic.se/support/downloads>


 Boule Medical AB
 Domnarvsgatan 4
 SE-163 53 Sp nga, Sweden

25122-3

For CA620 with barcode reader

1. Go to Menu 7.4.1

<p>CA620 Main Menu Program: V3c75</p> <ol style="list-style-type: none"> 1. Sample memory... 2. Set auto-print-mode=0 - 3. Set print-key function=2 - 4. Set print-format= 8 - 5. Setup menu... 6. Service menu... 7. QC menu... 8. System flow menu... 	<p>7. QC menu</p> <ol style="list-style-type: none"> 7.1 Control memory 7.2 Control L-J plots 7.3 X-B L-J plots 7.4 QC setup menu... 7.5 Set SEQ= 1 	<p>7.4.1 Input control definition</p> <p>Use the barcode reader to input the list of barcodes from the assay sheet for the control you want to define. See User's Manual for further info.</p> <p>0 barcodes read. Press <Menu> to skip.</p>
--	---	--

2. Scan in the calibrator/control in sequence top to bottom

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-51
ID: 3180351+
Exp.: 9-Aug-2018

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-52
ID: 3180352+
Exp.: 7-Aug-2018

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-53
ID: 3180353+
Exp.: 7-Aug-2018

1:	 Boul eCBAV1	 Boul eCBAV1	 Boul eCBAV1
2:	 3180351+	 3180352+	 3180353+
3:	 13180351	 23180352	 33180353
4:	 18080900001153010000	 18080700002153010000	 18080700003153010000
5:	 21712659700550314330	 40318798701160432235	 47622872701490541540
6:	 09515096152532038540	 27430098152882036040	 58560102153132035940
7:	 03403016040020301608	 09306047060060504010	 22218143150131006625
8:	 13245467700575047670	 12445508500655042750	 12445644500605029650
9:	 99994791588708472595	 99991456504574370351	 99990252593124253461

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 31803-51
 Prod.: Con-Diff Low
 ID: 3180351+
 Exp.: 9-Aug-2018

Instr.: M-series
 Lot: 31803-52
 Prod.: Con-Diff Norm
 ID: 3180352+
 Exp.: 7-Aug-2018

Instr.: M-series
 Lot: 31803-53
 Prod.: Con-Diff High
 ID: 3180353+
 Exp.: 7-Aug-2018

