

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹



Control	Lot	
Tri-Level	31803-8K	2018-08-07
Low	31803-51	2018-08-09
Normal	31803-54	2018-08-09
High	31803-53	2018-08-07

MEDONIC

CA620/530

		CONTROL	L
		LOT	31803-51
		Mean ²	± Range ³
RBC	10 ¹² /L	2,17	0,12
MCV	fL	65,9	7,0
RDW	%	13,2	4,5
Hct	%	14,3	3,0
Plt	10 ⁹ /L	95	15
MPV	fL	9,6	1,5
WBC	10 ⁹ /L	3,4	0,3
Hgb	g/dL	5,5	0,3
MCH	pg	25,3	2,0
MCHC	g/dL	38,5	4,0
Lymph#	10 ⁹ /L	1,6	0,4
Mid#	10 ⁹ /L	0,2	0,3
Gran#	10 ⁹ /L	1,6	0,8
Lymph	%	46,7	7,0
Mid	%	5,7	5,0
Gran	%	47,6	7,0

		CONTROL	N
		LOT	31803-54
		Mean ²	± Range ³
		4,01	0,18
		81,1	7,0
		12,3	4,5
		32,5	3,5
		264	30
		9,8	1,5
		9,1	0,6
		11,7	0,4
		29,2	2,0
		36,0	4,0
		4,6	0,6
		0,6	0,5
		3,9	1,0
		50,5	5,0
		6,8	5,0
		42,7	5,0

		CONTROL	H
		LOT	31803-53
		Mean ²	± Range ³
		4,76	0,22
		87,2	7,0
		12,4	4,5
		41,5	4,0
		585	60
		10,2	1,5
		22,2	1,8
		14,9	0,5
		31,3	2,0
		35,9	4,0
		14,3	1,5
		1,3	1,0
		6,6	2,5
		64,4	5,0
		6,0	5,0
		29,6	5,0

		Mean ²	± Range ³
RBC	10 ¹² /L	2,16	0,12
MCV	fL	67,4	7,0
RDW	%	13,2	4,5
Hct	%	14,6	3,0
Plt	10 ⁹ /L	85	15
MPV	fL	9,8	1,5
WBC	10 ⁹ /L	3,3	0,3
Hgb	g/dL	5,4	0,3
MCH	pg	25,0	2,0
MCHC	g/dL	37,0	4,0
Lymph#	10 ⁹ /L	1,5	0,4
Mid#	10 ⁹ /L	0,2	0,3
Gran#	10 ⁹ /L	1,6	0,8
Lymph	%	45,4	7,0
Mid	%	5,4	5,0
Gran	%	49,2	7,0

		Mean ²	± Range ³
		4,02	0,18
		83,5	7,0
		12,2	4,5
		33,6	3,5
		241	30
		9,8	1,5
		8,8	0,6
		11,8	0,4
		29,4	2,0
		35,1	4,0
		4,3	0,6
		0,6	0,5
		3,9	1,0
		48,8	5,0
		6,4	5,0
		44,8	5,0

		Mean ²	± Range ³
		4,82	0,22
		90,6	7,0
		12,1	4,5
		43,7	4,0
		536	60
		10,0	1,5
		21,5	1,8
		15,2	0,5
		31,5	2,0
		34,8	4,0
		13,1	1,5
		1,5	1,0
		6,9	2,5
		61,0	5,0
		7,1	5,0
		31,9	5,0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dana (hr)
 - cs Stabilitnosta otvorené ampulky 14 dní (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jours (fr)
 - de Stabilit t ge ffneter Flaschen 14 Tage (de)
 - el Διατ ρηση δειγματος μετ  την αποσφραγίση 14 ημερ  (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no  pen r r/glass stabilitet 14 dager (no)
 - pl Trwało c po otwarciu fiolek 14 Liczba dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekcnej liekovky 14 dni (sk)
 - sl Stabilita otvorene injekcne liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 d as (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Средно
 - Sredstvo
 - Pr m r
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - M so
 - M dia
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Średnia
 - M dia
 - M die
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povpre no
 - M dia
 - Medelv rde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο ε ρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Aralıđı



<http://www.medonic.se/support/downloads>


 Boule Medical AB
 Domnarvsgatan 4
 SE-163 53 Sp nga, Sweden

25122-3

For CA620 with barcode reader

1. Go to Menu 7.4.1























<p>CA620 Main Menu Program: V3c75</p> <ol style="list-style-type: none"> 1. Sample memory... 2. Set auto-print-mode=0 - 3. Set print-key function=2 - 4. Set print-format= 8 - 5. Setup menu... 6. Service menu... 7. QC menu... 8. System flow menu... 	<p>7. QC menu</p> <ol style="list-style-type: none"> 7.1 Control memory 7.2 Control L-J plots 7.3 X-B L-J plots 7.4 QC setup menu... 7.5 Set SEQ= 1 	<p>7.4.1 Input control definition</p> <p>Use the barcode reader to input the list of barcodes from the assay sheet for the control you want to define. See User's Manual for further info.</p> <p>0 barcodes read. Press <Menu> to skip.</p>
--	---	--

2. Scan in the calibrator/control in sequence top to bottom

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-51
ID: 3180351+
Exp.: 9-Aug-2018

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-54
ID: 3180354+
Exp.: 9-Aug-2018

Mfg: CA Boule reag.
Product: Con-Diff
Lot: 31803-53
ID: 3180353+
Exp.: 7-Aug-2018

<p>1:  Boul eCBAV1</p>	<p> Boul eCBAV1</p>	<p> Boul eCBAV1</p>
<p>2:  3180351+</p>	<p> 3180354+</p>	<p> 3180353+</p>
<p>3:  13180351</p>	<p> 23180354</p>	<p> 33180353</p>
<p>4:  18080900001153010000</p>	<p> 18080900002153010000</p>	<p> 18080700003153010000</p>
<p>5:  21712659700550314330</p>	<p> 40118811701170432535</p>	<p> 47622872701490541540</p>
<p>6:  09515096152532038540</p>	<p> 26430098152922036040</p>	<p> 58560102153132035940</p>
<p>7:  03403016040020301608</p>	<p> 09106046060060503910</p>	<p> 22218143150131006625</p>
<p>8:  13245467700575047670</p>	<p> 12345505500685042750</p>	<p> 12445644500605029650</p>
<p>9:  99994791588708472595</p>	<p> 99995568496237406462</p>	<p> 99990252593124253461</p>

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 31803-51
 Prod.: Con-Diff Low
 ID: 3180351+
 Exp.: 9-Aug-2018

Instr.: M-series
 Lot: 31803-54
 Prod.: Con-Diff Norm
 ID: 3180354+
 Exp.: 9-Aug-2018

Instr.: M-series
 Lot: 31803-53
 Prod.: Con-Diff High
 ID: 3180353+
 Exp.: 7-Aug-2018

